



Colloidal Oatmeal System

Nitrile Exam Gloves Powder Free, Standard Cuff

COATS® (an acronym for colloidal oatmeal system) is a patented and unique nitrile glove technology. COATS® utilises the powerful benefits of all-natural oats, an FDA-recognised skin protectant, as a coating that forms a natural, moisturising barrier between the glove and skin. This acts as a preventative measure against skin irritation, and eliminates many of the uncomfortable and irritating conditions experienced when wearing normal gloves. Users who suffer from dry and itchy skin due to constant hand washing and glove usage can now rely on COATS® to soothe and nurture the skin, and protect their hands while they work.



	COATS	® Nitrile	
Length (mm)			
	≥ 2	30	
Thickness Measurements (mm)			
Palm (centre of Palm)	0.07	<u>+</u> 0.02	
Finger (13mm \pm 3mm from tip)	0.09 <u>+</u> 0.02		
Physical Properties	Before Ageing	After Ageing	
Tensile Strength (MPa)	≥ 18	≥ 16	
Elongation (%)	≥ 500	≥ 400	
Inspection Levels & AQL	Inspection Level	AQL	
Watertightness	G1	1.5	
Physical Dimensions	S2	4.0	
Physical Properties	S2	4.0	
Visual Inspection (Major)	S4	2.5	
Visual Inspection (Minor)	S4	4.0	
Particulate Residue	N = 5	≤ 2mg/glove	
Colloidal Oatmeal Content	N = 5	≥ 5mg/glove	

Chemotherapy Drugs and Concentration (Tested for Resistance to Permeation by Chemotherapy Drugs as per ASTM D6978-05)	Minimum Breakthrough Detection Time (minutes)
Carmustine (BCNU), 3.3mg/ml (3,300 ppm)	Not recommended
Cisplatin, 1.0mg/ml (1,000 ppm)	>240 minutes
Cyclophosphamide (Cytoxan), 20.0mg/ml (20,000 ppm)	>240 minutes
Dacarbazine (DTIC), 10.0mg/ml (10,000 ppm)	>240 minutes
Doxorubicin Hydrochloride, 2.0mg/ml (2,000 ppm)	>240 minutes
Etoposide (Toposar), 20.00mg/ml (20,000 ppm)	>240 minutes
Fluorouracil, 50.0mg/ml (50,000 ppm)	>240 minutes
Methotrexate, 25.0mg/ml (25,000 ppm)	>240 minutes
Mitomycin C, 0.5mg/ml (500 ppm)	>240 minutes
Paclitaxel (Taxol), 6.0mg/ml (6,000 ppm)	>240 minutes
Thiotepa, 10.0mg/ml (10,000 ppm)	Not recommended
Vincristine Sulfate, 1.0mg/ml (1,000 ppm)	>240 minutes

WARNING: Gloves used for protection against chemotherapy drug exposure should be selected specifically for the type of chemicals being used. Due to the variety and concentration of chemotherapy drugs used in treatments, the resistance table shown does neither warrant nor imply the safe use of the gloves against chemotherapy drugs resistance in every case. The safe use of gloves in chemotherapytreatment is solely the decision of clinicians authorised to make such decision.

FEATURES

- Fingertip textured
- Powder free
- Not made with natural rubber latex
- Chemo drugs tested
- Lab chemical tested
- Ambidextrous
- Standard cuff
- Dawn blue colour

PACKAGING

100 gloves per box (XS-L) 90 gloves per box (XL) 10 boxes per carton

REGULATORY COMPLIANCE

TGA - ARTG 164563, FDA 510(k), MDD 93/42/EEC, REACH, EC 10/2011, EC 1935/2004

STANDARDS

ASTM D6319, ASTM D412, ASTM D573, ASTM D5151, ASTM D6124, EN 455 part 1, 2, 3 & 4, EN 1186, EN 13130, CEN/TS 14234

PATENTS

Patent 7,691,436; Patent 7,718,240; Patent 7,740,622; Patent 8,075,965; Patent 8,458,818

MANUFACTURING ACCREDITATIONS

ISO 9001 ISO 13485 EN ISO 13485 ISO 14001 OHSAS 18001

COATS® Colloidal Oatmeal Coated Nitrile Powder Free 2.5 Mil

ASTM D3578

Physical Dimensions			
Glove Length (mm)	≥ 230		
Palm Thickness (mm)	0.07 ± 0.02		
Finger Thickness (mm)	0.09 ± 0.02		
Physical Properties			
Test	Before Aging	After Aging	
Tensile strength (MPa)	≥ 18.0	≥ 16.0	
Elongation (%)	≥ 500	≥ 400	

EN 455

Physical Dimensions			
Median glove length (mm)	≥ 2	40	
Median palm thickness (mm)	0.07 ±	0.02	
Median finger thickness (mm)	0.09 ±	0.02	
Physical Properties			
Test	Before Aging	After Aging	
Median Force at break (N)	≥ 6	≥ 6	



Regulatory Compliance

FDA 510(k), MDD 93/42/EEC, REACH, ROHS Directive 2002/95/EC, EC 10/2011, EC 1935/2004, PPE 89/686/EEC

Standards

ASTM D6319, ASTM 6978, EN455 part 1, 2, 3 & 4, EN 1186, EN 13130, CEN/TS 14234, EN 420, EN 374 part 1, 2 & 3

Classification

Class I (FDA), Class I (MDD 93/42/EEC), Category 3 (BfR XXI), Category III (PPE 89/686/EEC)

Patent

7,691,436; 7,718,240; 7,740,622; 8,075,965; 8,458,818

Application Settings

Low risk - medical, dental, procedures, chemotherapy drugs, pathology lab and food handling. Coated with FDA recognised skin protectant. Clinically proven to help protect and moisturise your skin from dry and irritated skin from prolonged glove use and hand wash.

Colour

Dawn blue, white

MATERIAL SAFETY DATA SHEET





COMMON NAME (USED ON LABEL)
Nitrile Powder Free Examination Gloves

CHEMICAL FAMILY

Carboxylated Butadiene Acrylonitrile Polymer Latex

APPLICATION

TRADENAME & SYNONYM GLOVEON COATS NITRILE (CTS38)

NITRILE POWDER FREE EXAMINATION GLOVES COATS

Medical and Dental

SECTION 2: HAZARDOUS INGREDIENTS

HAZARDOUS COMPONENT	CAS #	%(WT)	TLV	PEL	
N/A	N/A	N/A	N/A	N/A	

PEL: Permissible Exposure Limit established by Occupational Safety and Health Administration (OSHA).

TLV: Threshold Limit Value established by the American Conference of Governmental Industrial Hygienists, 1987-1988.

SECTION 3: COMPOSITION/INFORMATION ON INGREDIENTS

CHEMICAL COMPOSITION

All chemicals used are non-toxic/ non-hazardous.

Butadiene-Acrylonitrile Latex, Sodium Dodecylbenzenesulfonate, Sulphur, Zinc Oxide, Zinc Di-nbutyldithiocarbamate, Titanium Dioxide, Paraffin Wax Emulsion

Coating Ingredient

Colloidal Oatmeal & Constituents, Sodium Benzoate, Processing Aid

SECTION 4: FIRST AID MEASURE

If reaction in the form of skin irritation is noticed, remove gloves immediately and wash affected part with saline water. If there is no relief, seek medical reactions.

SECTION 5: FIRE FIGHTING MEASURE

FLASHPOINT	AUTOIGNITION TEMPERATURE	FLAMMABLE LIMITS IN AIR
N/A	N/A	N/A

EXTINGUISHING MEDIA

Chemical foam and dry chemical may be used.

FIRE-FIGHTING PROCEDURES

Use standard procedures for combustion material fires, including approved self-contained breathing apparatus.

FIRE AND EXPLOSION HAZARDS

No fire or explosion hazards are associated with these products. They will melt at elevated temperatures.

SECTION 6: ACCIDENTAL RELEASE MEASURES

BIOCOMPATABILITY

The chemical formulation of the gloves and surface lubrication materials does not contain any substances normally known to be harmful to the user or to any person with whom the gloves come into contact.

MEDICAL CONDITIONS GENERALLY AGGRAVATED BY EXPOSURE

Nitrile Powder Free Gloves are not expected to cause any adverse health effects.

SECTION 7: HANDLING AND STORAGE

PRECAUTIONS TO BE TAKEN IN HANDLING AND STORAGE

Store in a dry, cool and ventilated area. Do not store above 104 $^{\circ}$ F (40 $^{\circ}$ C). Shield open box from direct sunlight, fluorescent lighting and x-rays. Improper storage will decrease usable life.

SECTION 8: EXPOSURE CONTROLS/ PERSONAL PROTECTION		
EYE PROTECTION Not necessary under conditions of intended use.	SKIN PROTECTION Not necessary under conditions of intended use.	
RESPIRATORY PROTECTION Not necessary under conditions of intended use.	VENTILATION Not necessary under conditions of intended use.	

STEPS TO BE TAKEN IN CASE MATERIAL IS LEAKED OR SPILLED

These products are solid articles and are not subject to leaks or spills.

SECTION 9: PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE/ ODOR

Ambidextrous, Beaded Cuff, Micro-textured, Chlorinated, Powder Free, Coated with Colloidal Oatmeal USP Skin Protectant, Dawn Blue.

DIMENSION	X-SMALL	SMALL	MEDIUM	LARGE	X-LARGE		
Length (mm)		Mi	nimum 230 (same fo	r all)	e e		
Width (mm)	76 ± 4	86 ± 4 98 ± 4 107 ± 4 115					
THICKNESS (mm	- SINGLE WALL	MEASUREMENT (same for all)				

Finger (mm) 0.09 ± 0.02 Palm (mm) 0.07 ± 0.02

TENSILE PROPERTIES	UNAGED	AGED
Tensile Strength (Mpa)	Min. 18.0 MPa	Min. 16.0 MPa
Ultimate Elongation (%)	Min. 500%	Min. 400%

SECTION 10: STABILITY AND REACTIVITY

BOILING POINT	VAPOR PRESSURE (mm Hg)	VAPOR DENSITY (air=1)
N/A	N/A	N/A
SPECIFIC GRAVITY (water=1) N/A	SOLUBILITY IN WATER Insoluble	% VOLATILE BY VOLUME N/A

EVAPORATION RATE VISCOSITY
N/A N/A

SECTION 11: TOXICOLOGICAL INFORMATION

STABILITY
Stable.

CONDITIONS TO AVOID
Does not apply.

INCOMPATABILITY (MATERIALS TO AVOID)

High polar solvent like methyl ethyl ketone, acetone.

HAZARDOUS DECOMPOSITION PRODUCTS

In a fire, these products may produce a black smoke. Carbon Dioxide, Carbon Monoxide, Oxides of Nitrogen, aromatic/aliphatic hydrocarbons.

HAZARDOUS POLYMERIZATION

Will not occur.

SECTION 12: ECOLOGICAL INFORMATION

N/A

SECTION 13: DISPOSAL CONSIDERATION

WASTE DISPOSAL METHOD

Consult current local, state and federal regulations for proper disposal methods.

SECTION 14: TRANSPORT INFORMATION

N/A

SECTION 15: REGULATORY INFORMATION

N/A

SECTION 16: OTHER INFORMATION

RECOMMENDED USE AND RESTRICTION

The Nitrile Powder Free Gloves is a Single Use device.

The Brand

Leadership | Certifications | Global Locations

Certifications

Certifications

Gloveon's quality standards, management systems and exemplary regulatory compliance, all contribute to the global success of the company. Our capabilities have been assessed and certified by the following international governing bodies.



Management Service ISO 9001:2015



America ISO 13485:2016



EN ISO 13485:2016



Confirmation Letter for GMP Audit



EC Certificate



ISO 14001:201



UL Certification



ISEGA Food Contact Test Certification (German)



Registration Certificate for Medical Device



NFPA Certification



510(k) Approval



PPE Cert









EC Declaration of Conformity

We, the manufacturer Hartalega Sdn. Bhd., No. 7, Kawasan Perusahaan Suria, 45600 Bestari Jaya, Selangor Darul Ehsan, Malaysia

is in conformity with the relevant Union harmonisation legislation PPE Regulation (EU) 2016/425 $\,$

where such is the case, with the national standard transposing harmonized standard number EN 420: 2003+A1:2009 EN ISO 374-1:2016 EN ISO 374-5:2016

The notified body SATRA Technology Europe with Notified Body Number of 2777 performed the EU type-examination (Module B) and issued the EU type-examination certificate 2777/10475-03/E00-00.

the PPE is subject to the conformity assessment procedure conformity to type based on internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the Notified body SATRA Technology Europe with Notified Body Number of 2777.

Done at Hartalega Sdn. Bhd. on 11th February 2020.

AMAIN

Kuan Eu Jin
Quality Management Representative



EC Declaration of Conformity

We, the manufacturer Hartalega Sdn. Bhd., No. 7, Kawasan Perusahaan Suria, 45600 Bestari Jaya, Selangor Darul Ehsan, Malaysia

with European Representative Medical Device Safety Service (MDSS) Schiffgraben 41, 30175 Hannover, Germany

Declares that the new PPE described hereafter
Category III (Type 8)
HSB-TT-005
2.2 S mil Powder Free Nitrile disposable five fingered glove
Available in a standard minimum 240mm length or a longer cuff variant of 280mm
Available in sterile and non-sterile

is in conformity with the relevant Union harmonisation legislation PPE Regulation (EU) 2016/425

where such is the case, with the national standard transposing harmonized standard number

The notified body SATRA Technology Centre with Notified Body Number of 2777 performed the EU type-examination (Module B) and issued the EU type-examination certificate 2777/11755-02/E00-00.

The PPE is subject to the conformity assessment procedure conformity to type based on internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the Notified body SATRA Technology Centre with Notified Body Number of 2777.

Done at Hartalega Sdn. Bhd. on 11th February 2020.

Done at Hartalega Sdn. Bhd. on 11th F Kuan Eu Jin Quality Management Representative





EC Declaration of Conformity

We, the manufacturer Hartalega Sdn. Bhd., No. 7, Kawasan Perusahaan Suria, 45600 Bestari Jaya, Selangor Darul Ehsan, Malaysia

with European Representative Medical Device Safety Service (MDSS) Schiffgraben 41, 30175 Hannover, Germany

Declares that the new PPE described hereafter

Category III (Type B)
HSB-TF-009
Nitrile Powder Free Gloves with Colloidal Oatmeal USP Skin Protectant

where such is the case, with the national standard transposing harmonized standard number EN 420: 2003+A1: 2009 EN ISO 374-12016 EN ISO 374-52016

The notified body SATRA Technology Centre with Notified Body Number of 2777 performed the EU type-examination (Module B) and issued the EU type-examination certificate 2777/10783-02/E00-00.

The PPE is subject to the conformity assessment procedure conformity to type based on internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the Notified body SATRA Technology Centre with Notified Body Number of 2777.

Done at Hartalega Sdn. Bhd. on 11th February 2020.

Kuan Eu Jin Quality Management Representative





Hartalega Sdn Bhd Nurul Kong Quality Assurance Senior Mana No. 7, Kawasan Perusahaan Sur Bestari Jaya, 45600 My

Re: K180644
Trade/Device Name: Nitrile Powder Free Examination Gloves with Colloidal Oatmeal-Lemon Green
Regulation Number: 21 CFR 880 6250
Regulation Name: Patient Examination Glove
Regulationy Class: Class 1
Product Code: LZA Dated: July 16, 2018 Received: July 23, 2018

Dear Nurul Kong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predictate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmeic act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The \$10(k) Premarket Notification Database located at <a href="https://www.accesselian.tide.gov/scripts/edh/cfdc/sec/fpmip/mem.efm/lettrifice/combination/products/submissions.The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations. Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to registration and listing (21 CFR Part 807); helping (21 CFR Part 807); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for

U.S. Food & Drug Administration 10903 New Hampshire Avenue Silver Spring, MD 20993

devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see https://www.fda.gov/CombinationProducts/GidlanceRegulatory/information/ucm97-2488 hm; good manufacturing practice requirements as set forth in the quality systems (S05) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and papilicable, the electronic product radiation control provisions (Sections 315-454 of the Act.); 21 CFR 1000


Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807-97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to https://www.fds.gov/Meds/alb/cvcses/Saft/Report/Problem/default.htm.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fds.gow/MedicalPevices/Devices/Equalatoma/Giudance) and CDRH Leam (http://www.fds.gow/Training/CDRHLeam). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (http://www.fds.gow/DICE) for more information or contact DICE by email (DICE/Edfalahis.gov) or phone (1500-6438-2044) or 301-796-7100.

Clarence W. Murray lii III -S

For Tina Kiang, Ph.D.
Acting Director
Division of Anesthesiology,
General Hospital, Respiratory,
Infection Control, and Dental Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure





510(k) Premarket Notification

FDA Home
Medical Devices
Databases



510(k) | DeNovo | Registration & Listing | Adverse Events | Recalls | PMA | HDE | Classification | Standards CFR Title 21 | Radiation-Emitting Products | X-Ray Assembler | Medsun Reports | CLIA | TPLC

New Search Back To Search Results

Device

Classification

Polymer Patient Examination Glove

Name

510(K) Number K133956

Device Name

NITRILE POWDER FREE EXAMINATION GLOVE WITH COLLOIDAL OATMEAL USP SKIN PROTECTANT DRUG - WHITE / DAWN BLUE / LEMON GREEN

HARTALEGA SDN BHD **Applicant**

NO. 7, KAWASAN PERUSAHAAN SURIA

Bestari Jaya, Selangor, MY 45600

Applicant Nurul Aisyah Kong Contact

Correspondent HARTALEGA SDN BHD
NO. 7, KAWASAN PERUSAHAAN SURIA
Bestari Jaya, Selangor, MY 45600

Contact

Correspondent Nurul Aisyah Kong

Regulation 880.6250 Number Classification LZA

Product Code

Date Received 12/23/2013 Decision Date 05/28/2014

Decision Substantially Equivalent (SESE)

Regulation

Medical General Hospital Specialty

510k Review

General Hospital Panel

Summary Summary Traditional Type Reviewed By Third Party

Combination No **Product**

FDA

FDA Home³ Medical Devices⁴ Databases⁵

New Search ¹⁵ Help ¹⁶			Export To	Excel
Device Name		▲19 ▼20	510(K) Number +22	Decision Date
Powdered Sterile Latex Surgical Glove, With Protein Content Labeling Claim (200 Micrograms Or Less)	HARTALEGA SDN BHD		K001959	07/26/2000
Powder Free Sterile Latex Surgical Gloves, Contains 50 Microgram Or Less Of Total Water Extractable Protein Per Gram	HARTALEGA SDN BHD		K002593	11/29/2000
Freeform Blue Powderfree Nitrile Examination Gloves	HARTALEGA SDN BHD		K022671	11/18/2002
Freeform Blue Powder-free Nitrile Examination Gloves	HARTALEGA SDN BHD		K041391	07/09/2004
Nitrile Powder Free Examination Gloves (White)	HARTALEGA SDN BHD		K050214	03/16/2005
Nitrile Powdered Examination Gloves (White)	HARTALEGA SDN BHD		K050215	03/11/2005
Chlorinated Powder Free Latex Examination Gloves (Yellow)	HARTALEGA SDN BHD		K050277	06/07/2005
Nitrile Powder Free Examination Gloves (Blue)	HARTALEGA SDN BHD		K051777	08/12/2005

1 of 4

5/15/2020, 5:04 PM



Testing. Development. Problem Solving.

April 15, 2009

· TEST REPORT ·

PN 83672A - Amended

CHEMICAL ANALYTICAL SERVICES

Prepared For: Hartalega SDN. BDH Ms. Nurul Aisyah Kong No. 7 Kawasan Perusahaan Suria Bestari Jaya Selangor, 45600 Malaysia

edited Testing Laboratory — Certificate Numbers 255.01 & 255.02 ISO 6001:2000 Registered of ACIL: The American Council of Independent Laboratories

ISO 9001:2000



Testing. Development. Problem Solving.

April 15, 2009

Ms. Nurul Aisyah Kong Hartalega SDN. BHD

Page 1 of 3 - PN 83672A - Amended

<u>SUBJECT:</u> Permeation testing per ASTM D 6978-05 on sample submitted by the above company. Wire Transfer.

RECEIVED: Glove sample identified as Nitrile Powder Free Examination Gloves (Blue) Code: ABLU.

TESTING CHEMOTHERAPY DRUGS:

Table 1. List of the Testing Chemotherapy Drugs, Sources, and Expiration Dates

TESTING CHEMOTHERAPY DRUGS	DRUG SOURCE
Carmustine (BCNU)	Sigma, Lot# 038K4008; Expiration 12/2009
Cisplatin	Sigma; Lot# 59H3657; Expiration 09/2009
Cyclophosphamide (Cytoxan)	Sigma; Lot# 068K1131; Expiration 1/2010
Dacarbazine (DTIC)	Hospira; Lot# U022223AA; Expiration 06/2010
Doxorubicin Hydrochloride	Teva; Lot#07N625; Expiration 10/2009
Etoposide (Toposar)	Teva; Lot# 31303976B; Expiration 9/2011
Fluorouracil	APP; Lot# 203867; Expiration 03/2010
Mitomycin C	Sigma, Lot# 048K1086; Expiration 01/2010
Methotrexate	Hospira; Lot# U024457AA; Expiration 05/2010
Paclitaxel (Taxol)	Dabur Oncology; Lot# PA08H00701; Exp. 05/2010
Thiotepa	Sigma; Lot#078K1526; Expiration 12/2009
Vincristine Sulfate	Hospira, Lot# U037139AA; Expiration 12/2009

COLLECTION MEDIA: The collection media, which were selected, are listed in Table 2.

Table 2. Collection Media for Testing Chemotherapy Drugs

TEST DRUG AND CONCENTRATION	COLLECTION MEDIUM					
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	10% Ethanol Aqueous Solution					
Cisplatin, 1.0 mg/ml (1,000 ppm)	Distilled Water					
Cyclophosphamide (Cytoxan), 20 mg/ml (20,000 ppm)	Distilled Water					
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	Distilled Water					
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	Distilled Water					
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	Distilled Water					
Fluorouracil, 50.0 mg/ml (50,000 ppm)	9.20 pH Sodium Hydroxide Solution					
Methotrexate, 25 mg/ml (25,000 ppm)	Distilled Water					
Mitomycin C, 0.5 mg/ml (500 ppm)	Distilled Water					
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	30% Methanol Aqueous Solution					
Thiotepa, 10.0 mg/ml (10,000 ppm)	Distilled Water					
Vincristine Sulfate, 1.0 mg/ml (1,000 ppm)	Distilled Water					



MDSS - Schiffgroben 41 - 30175 Hannover, Germany

Hartalega NGC Sdn Bhd.
Khairunnisa Warsito
No. 1, Persiaran Tanjung
Kawasan Perindustrian Tanjung
43900 Sepang, Selangor
MALAYSIA

Schiffgroben 41 30175 Hannover, Germany

Tel: + 49 - 511 - 62 62 86 30 Fox: + 49 - 511 - 62 62 86 33

eMail: info@mdss.com Internet: www.mdss.com

2019.01.18

Confirmation of CE Registration

Dear Khairunnisa

It is our pleasure to enclose the new Certificate of CE-Registration for your product.

Please note that registration was performed under § 25 MPG (Medizinproduktegesetz). This is the Federal Republic of Germany's national interpretation of Medical Device Directive 93/42/EEC. Registration is therefore in accordance with EU legislation. We remind you that all products must meet the applicable provision of the European and national regulation before they may be placed on the market.

We are looking forward to continuing our good business relationship and wish you a successful product launch in Europe.

Best regards,

Encl. 1 Certificate of CE-Registration 1 Annex A

MDSS - Medical Device Safety Service GmbH Handelsregister Hannover HRB 57318 - USt-IdNr. DE 177346163 - Geschäftsführer: Ludger Möller

Handelsregister Transition
Bankverbindungen
Sparkasse Honnover
S.W.I.F.T.: SPHHDE2H
IBAN: DE24 2505 0180 0910 0792 77

Commerzbank AG, Hannover S.W.I.F.T.: COBADEFF 250 IBAN: D667 2504 0066 0338 8816 00







Certificate of CE-Registration

This is to certify that, in accordance with the Medical Device Directive 93/42/EEC, Medical Device Safety Service GmbH (MDSS) agrees to perform all duties and responsibilities as the Authorized Representative for:

Hartalega NGC Sdn. Bhd. No. 1, Persiaran Tanjung Kawasan Perindustrian Tanjung 43900 Sepang, Selangor MALAYSIA

as stipulated and demanded by the aforementioned Directive. The German Competent Authority has allocated the medical devices of the Manufacturer registration numbers as foreseen in:

Annex A dated January 18, 2019

The Manufacturer has provided MDSS with the appropriate Declaration(s) of Conformity confirming that the medical devices fulfill the applicable requirements of Directive 93/42/EEC. In compliance with German law, a safety officer has been appointed for Germany.

2019-01-18

Ludger Möller President MDSS GmbH

 $\textbf{MDSS} \cdot \text{Medical Device Safety Service} \cdot \text{Schiffgraben 41} \cdot 30175 \ \text{Hannover, Germany}$



April 25, 2020

Hartalega NGC SDN. BHD. Nurul Kong Senior Manager- Quality Assurance Kawasan Perindustrian Tanjung Sepang, Selangor 43900 Malaysia

Re: K200581

K200581
Trade/Device Name: Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (Blue)
Regulation Number: 21 CFR 880 6250
Regulation Name: Non-Powdered Patient Examination Glove
Regulator (Sais: Class I, reserved
Product Code: LZA, LZC, QDO
Dated: February 27, 2020
Received: March 5, 2020

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at https://www.accessdata.ida.gov/scripts/edits/effocss/c/pmn/pmn_cfm_dentifice_combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal

U.S. Food & Drug Administration 10903 New Hampshire Avenue Silver Spring, MD 20993

statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 808)) medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmatering asfelv promping (21 CFR 4, 8) alphan Bf) or combination products (see <a href="https://www.fda.gov/combination-products/guidance-regulatory-information-postmarketing-safely-reporting-combination-products); good manufacturing practice regulatory-information postmarketing-safely-reporting-combination-products; good manufacturing practice regulatory-information products (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <a href="https://www.fda.gov/medical-device-safety/medical-device-reporting-medi-hou-report-nedical-device-report-nedical-devic

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance) and CDRH Learn (https://www.fda.gov/traning-and-continuing-aducation/cdrh-learn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ake a question about a specific regulatory topic. See the DICE website (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-un-division-industry-and-consumer-decudation-dice) for more information or contact DICE by email (DICEgifda.hts.gov) or phone (1-800-638-2041 or 301-796-7100).

Elizabeth F. Claverie -S

Clavene -5
CAPT Elizabeh Claverie, M.S.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4- Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Form Approved; OMB No. 0910-0120 Expiration Date: 06/30/2020 See PRA Statement below. DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration Indications for Use 510(k) Number (if known) Device Name Biodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl Citrate (Blue) indications for Use (Describo)
Blodegradable Nitrile Powder Free Examination Gloves Tested for Use with Chemotherapy Drugs and Fentanyl, Citrate
(Blue) is a non-settle disposable device intended for medical purpose that is worn on the examiner's hand to prevent
contamination between patient and examiner. It is also tested to be used against Chemotherapy Drugs and Fentanyl Citrate. These gloves were tested for use with chemotherapy drugs and Fentanyl Citrate as per ASTM D6978-05 (Reapproved 2013) Standard Practice for Assessment of Medical Gloves to Permention by Chemotherapy Drugs. Cb

hemotherapy Drug and Concentration	Minimum Breakthrough Detection Time in Minute
Carmustine (3.3 mg/ml)	21.4
Cisplatin (1.0 mg/ml)	>240
Cyclophosphamide (20.0 mg/ml)	>240
Dacarbazine (10,0 mg/ml)	>240
Doxorubicin Hydrochloride (2.0 mg/ml)	>240
Etoposide (20.0 mg/ml)	>240
Pluorouracil (50.0 mg/ml)	>240
Methotrexate (25.0 mg/ml)	>240
Mitomycin C (0.5 mg/ml)	>240
Paclitaxel (6.0 mg/ml)	>240
Thiotepa (10.0 mg/ml)	67.2
Vincristine Sulfate (1.0 mg/ml)	>240
Azacytidine (25,0 mg/ml)	>240
Carboplatin (10.0 mg/ml)	>240
Docetaxel (10 mg/ml)	>240
Epirubicin (2.0 mg/ml)	>240
Gemcitabine (38 mg/ml)	>240
Ifosfamide (50 mg/ml)	>240
Irinotecan (20 mg/ml)	>240
Mitoxantrone (2.0 mg/ml)	>240
Oncovin (1.0 mg/inl)	>240
Oxaliplatin (5 mg/ml)	>240
Vinorelbine (10 mg/ml)	>240

Please note that Cannustine and Thiotopa have extremely low permeation times of 21.4 minutes and 67.2 minutes respectively.

Warning: Do not use with Carmustine

Minimum Breakthrough Detection Time in Minutes >240

Fentanyl Citrate and Concentration entanyl Citrate Injection (100 mcg/2ml) ise (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D) | X Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

FORM FDA 3881 (7/17) Page 1 of 2 FORM FDA 3881 (7/17)

This section applies only to requirements of the Paperwork Reduction Act of 1995.
*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

TUN NO 3 SAID TOUR COMPLETED FORM TO THE PRA STAP EMAIL ADDRESS BELLOW:
The burden time of the localedion of information is estimated to average 79 hours per response, including time to review instructions, search existing data sources, gather and maintain the data needed and compa and review the collection of information. Send comments registeding this turion estimate or any other arg of this information collection, including suggestions for reducing this burdon estimate or any other arg of this information collection, including suggestions for reducing this burdon. In the property of the property

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Hartalega Attains International Certification on Occupational Health and Safety – OHSAS 18001




Hartalega has once again proven its commitment to the highest quality standards, as the Group recently attained OHSAS 18001:2007 certification.


Awarded by TUV SUD Asia Pacific TUV SUD Group, an audit and management systems certification body, OHSAS 18001:2007 is an internationally recognised standard which sets the requirements and best practices for occupational health and safety management systems in an organisation. The Group was previously awarded ISO 14001:2004 certification as a result of its outstanding environmental management system.

Mr Kuan Mun Leong, Managing Director of Hartalega said, "The OHSAS is a testament to our group's commitment to the well being of all Hartanians. As we continue to grow our business aggressively, being able to provide a quality work place in the aspects of health and safety is very important."

The OHSAS 18001:2007 certification was achieved through Hartalega's comprehensive range of health and safety measures, which include internal workplace audits, risk assessments, behaviour observations, accident and incident investigations, work permit issuances, training sessions for emergency preparedness and environmental performance monitoring, amongst others.

"As important as it is to focus on productivity and efficiency, it is equally as crucial to ensure that our employees work in a safe environment. We aim to continuously enhance Health, Safety and Environment initiatives throughout the Group for the benefit of our workforce," concluded Kuan.









CRS REF: SAT/18/0248 DATE RECEIVED: MAR 02, 2018 DATE REPORTED: MAR 14, 2018 PAGE: 1 of 1

Report No. : CRSSA/02645/18

TEST REPORT

Product Description Country of Origin Size Quantity Tested Test Conducted Test Method Testing Period Powder Free Nitrile Examination Gloves Malaysia Medium 200 pieces Freedom from holes ErM45 Part 1:2000 02 Mar 2018 – 08 Mar 2018

Based on submitted samples, the following results obtained :-

: Within AQL Result

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.

Thee. CHEE TUCK CHOON B.Sc. MMIC SECTION HEAD

This document is issued by the Company subject to the General Conditions of Service protect owned, available on request or accessible as in Service and Conditions and Cond

SSG (Malaysig) 5dh. Bhd.
(Company No. 10871-T)

Lot 4, Persianna or vere nuceria evy.

Seksyen 22, 40300 Shah Alam, Selangor, Malaysia

£ +6 (03) 5481 8282 f +6 (03) 5481 8215 www.sgs.com

SGS

CRS REF: SAT/18/0248 DATE RECEIVED: MAR 02, 2018 DATE REPORTED: MAR 14, 2018 PAGE: 1 of 1

Report No. : CRSSA/02646/18

TEST REPORT

Powder Free Nitrile Examination Gloves Malaysia Medium 13 pieces Dimensions EN 455 Part 2:2015 02 Mar 2018 – 08 Mar 2018 Product Description Country of Origin Size Quantity Tested Test Conducted Test Method Testing Period

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width	98	98	96	98	98	97	98	97	98	97	96	97	97	97
Length	250	255	250	255	251	250	252	252	250	254	252	253	252	252

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.

Chee CHEE TUCK CHOON B.Sc. MMIC SECTION HEAD

SGS (Malaysia) Sdn. Bhd. Lot 4, Persiaran Jubli Perak, Seksyen 22, 40300 Shsh Alam, Selangor, Malaysia t+6 (03) 5481 8282 f+6 (03) 5481 8215 www.sgs.com



CRS REF: SAT/18/0248 DATE RECEIVED: MAR 02, 2018 DATE REPORTED: MAR 14, 2018 PAGE: 1 of 1

Report No. : CRSSA/02647/18

TEST REPORT

Product Description Country of Origin Size Quantity Tested Test Conducted Test Method Ageing IESJ REPORT.

Powder Free Nitrile Examination Gloves
Malaysia
Medium
13 pieces
Force at Brack During Shelf Life and After
Challenge EN 455 Part 2-2015
70 ± 2 Deg C for 168 hrs
02 Mar 2018 – 14 Mar 2018 Ageing Testing Period

		Force at Break, N						
SIZE	SAMPLE NO.	BEFORE AGING	AFTER AGING					
M	1	8.2	8.4					
	2	8.1	8.2					
	3	7.9	6.5					
	4	7.3	7.9					
	5	8.5	6.6					
	6	9.2	9.3					
	7	8.7	7.2					
	8	8.8	7.4					
	9	9.3	7.1					
	10	8.0	7.9					
	11	9.2	7.3					
	12	6.3	7.1					
	13	8.1	7.1					
Median		8.2	7.3					
Dogwiromont		>60	>60					

Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.

Chee. CHEE TUCK CHOON B.Sc. MMIC SECTION HEAD

SGS (Malaysia) SAn. Bhd. (L014, Persianan Jubi Perak, Seksyen 22, 40300 Shah Alam, Selangor, Malaysia (Company No. 10871-1) t+6 (03) 5481 8282 f+6 (03) 5481 8215 www.sgs.com



CRS REF: SAT/18/0248 DATE RECEIVED: MAR 02, 2018 DATE REPORTED: MAR 14, 2018 PAGE: 1 of 1

Report No. : CRSSA/02648/18

TEST REPORT

Powder Free Nitrile Examination Gloves Malaysia Medium 5 pieces Powder Content EN455 Part 3:2015 02 Mar 2018 – 08 Mar 2018 Product Description Country of Origin Size Quantity Tested Test Conducted Test Method Testing Period

On testing the samples, the following results were obtained:-

SIZE Average Powder Mass per Glove

Note: Upon Customer's request, this report has been issued in more than one original. Only the first original is a legally binding document and may be used for any legal purpose, including payment. (Original 1-3)

SGS (MALAYSIA) SDN. BHD.

Thee. CHEE TUCK CHOON B.Sc. MMIC SECTION HEAD

As transacts for wave any generated socie surgery or trace scene copy.

SGS (Malaysia) Sdn. Bhd.
(Company No. 16871-T)

4+6 (03) 5481 8282 f +6 (03) 5481 8215 www.sgs.com

25cm X12.5cm X 5cm. 100 Gloves in 1 box













26cm X 26cm X 26cm. 10 boxes of 100 Gloves in One Carton







